

KM



KEEP MEDICAL


PRODUCT MANUAL

— Non Woven Face Mask Single Use

PRODUCT INFORMATION



**NON WOVEN
FACE MASK
SINGLE USE**



Product Name: NON WOVEN FACE MASK SINGLE USE

Specification: 175mm x 95mm

Material: spun-bond and melt-bond non-woven

Color: Solid

Production Date: refer to package

Shelf Life: 2 years

Content and Structure:

Main body made of 3-ply non-woven, ear loops and nose clip are fixed with ultra-sonic facility. This is un-sterilized product, do not re-use.

Storage:

Please always keep clean and in indoor circumstances with humidity < 80%, no corrosive air and in good ventilation. Avoid keeping in high temperature circumstances.

Range of Use:

For using of covering wearer's mouth, nose and jaw. General medical circumstance is applicable as for isolate the wearer and avoid cross infection.

PRODUCT INFORMATION

Way of Wearing

1. please make sure product is within shelf life and package is well sealed before using
2. three easy steps for wearing: cover the face, fix the ear loops, and adjust the nose clip
3. please dispose properly and following local disposal requirement for medical waste



CERTIFICATE OF COMPLIANCE

CERTIFICATE OF COMPLIANCE

This product has been inspected and tested and proved to be qualified for delivery in conformity with standard.

Product Name: NON WOVEN

FACE MASK SINGLE USE

Specification: 175mm x 95mm

Material: spun-bond and melt-bond non-woven

Color: solid

Production Date: refer to package

Shelf Life: 2 years

Inspector:



MADE IN CHINA

WAY OF PACKING



Plastic Bag(50Pcs)

Box (50Pcs/Box)

Size: 100mm*90mm*195mm

Carton (50Boxes/Carton) 2500Pcs

Size: 510mm*470mm*400mm

BUSINESS INFORMATION



MOQ: 100K

Payment Terms: T/T before shipment

Banking Info:

Account Name:

Account Number:

Bank Name:

Transportation and Express Info:

Pick-up in factory is recommended for massive orders

SF express with freight collect is the default option unless other options agreed prior

Company Info:

Company Name: DONGGUAN KEEP MEDICAL TECHNOLOGY CO., LTD.

Registration Number: 91441900MA54BFMHXA

Address: No.31 Mingyingyuan Road, Qinghu, Qishi Town, Dongguan City, Guangdong, China

QUALIFICATION INFORMATION



TEST REPORT (Chinese)

TEST REPORT (Chinese)



检验检测报告 (电子版)
防伪查询网址: www.gttc.net.cn
防伪码: VLGB-7068-24
No: 200045401
共3页 第1页


 

委托单位	东莞市凯医科技有限公司 地址: 广东省东莞市金石镇清湖民营园路31号				
样品信息	日常防护型口罩 100个				
检验性质	委托检测	样品受理/测试开始日期	2020-03-20	报告签发日期	2020-03-30
判定依据	GB/T 32610-2016 《日常防护型口罩技术规范》				
综合检验结论	—				
检验检测结果	检验检测项目	判定依据	判定		
	pH值	GB/T 32610-2016	符合		
	甲醛含量	GB/T 32610-2016	符合		
	外观要求[10个]	GB/T 32610-2016	符合		
	视野	GB/T 32610-2016	符合		
	吸气阻力	GB/T 32610-2016	符合		
	呼气阻力	GB/T 32610-2016	符合		
备注	本报告中检验检测项目均在相应标准规定的环境条件下进行(有注明的除外)。复印件、副本未重新加盖报告书确认章无效。本报告检验检测地址为广州市番禺区珠江路1号。				


签发: 刘圆 工程师

总部: 广州市番禺区珠江路1号
花都实验室: 广州市花都区狮岭镇旗岭河滨西路1号
电话: 020-61994598/61994599
电话: 020-37721161

1-1



样品图片 (电子版)
No: 200045401
共3页 第2页



总部: 广州市番禺区珠江路1号
花都实验室: 广州市花都区狮岭镇旗岭河滨西路1号
电话: 020-61994598/61994599
电话: 020-37721161

1-2

QUALIFICATION INFORMATION



FDA CERTIFICATION



Fiscal Year 2020
CERTIFICATION OF REGISTRATION

This certifies that:

DONGGUAN KEEP MEDICAL TECHNOLOGY CO.,LTD
No. 31, Qinghu Private Garden Road, Qishi Town, Dongguan City,
Guangdong Province, 523518, CHINA

has completed the FDA Establishment Registration and Device Listing with the US Food & Drug Administration, through

Shanghai JAT Enterprise Management Consultation Co., Ltd.

Owner Operator Number:10063594

Device Listing#:

Listing No	Code	Device Name	Proprietary Name
D379247	LYU	ACCESSORY, SURGICAL APPAREL	Disposable Protect Mask DB-001, DB-002

JAT will confirm that such registration remains effective upon request and presentation of this certificate until the end of the calendar year stated above, unless said registration is terminated after issuance of this certificate. JAT makes no other representations or warranties, nor does this certificate make any representations or warranties to any person or entity other than the named certificate holder, for whose sole benefit it is issued. This certificate does not denote endorsement or approval of the certificate-holder's device or establishment by the U.S. Food and Drug Administration. JAT assumes no liability to any person or entity in connection with the foregoing.

Pursuant to 21 CFR 807.39, "Registration of a device establishment or assignment of a registration number does not in any way denote approval of the establishment or its products. Any representation that creates an impression of official approval because of registration or possession of a registration number is misleading and constitutes misbranding." The U.S. Food and Drug Administration does not issue a certificate of registration, nor does the U.S. Food and Drug Administration recognize a certificate of registration, JAT is not affiliated with the U.S. Food and Drug Administration.





Chief engineer
Issued: 03/26/2020
Expiration Date: 12/31/2020

KEEP MEDICAL